

S-Trap™ micro spin column digestion protocol

Required equipment and materials:

- 1.7 mL sample tubes
- Single channel pipettors with corresponding tips (P2, P20, P200, P1000)
- Benchtop centrifuge
- pH meter with electrode
- Vortex mixer
- Heat block set to 55 °C
- Water bath or humidified incubator set to 47 °C
- SpeedVac or lyophilizer
- Vacuum manifold (optional; Supelco Visiprep SPE Vacuum Manifold is recommended)
- Sonicator, PIXUL or Covaris ultrasonicators; automation. Optional and recommended; see www.protifi.com/pixul and www.protifi.com/A200.

Required reagents and solutions:

- Protein sample (**1 µg – 100 µg**)
- pH reference solutions
- Tris(2-carboxyethyl)phosphine (**TCEP**)
- Methyl methanethiosulfonate (**MMTS**)
- Protease of choice (e.g. trypsin; 1 µg per 10 µg of sample)
- Stock of 1 M triethylammonium bicarbonate (**TEAB**), pH 8.5
- Aqueous phosphoric acid
- Sodium dodecyl sulfate (**SDS**)
- LC/MS grade methanol (**MeOH**), formic acid, acetonitrile, water and isopropanol
- Benzoylase and 50 mM MgCl₂ (optional)


Required solutions (available for separate purchase at www.protifi.com):

Solution	Composition	pH	Storage
1X lysis buffer	5% SDS, 50 mM TEAB pH 8.5; make by diluting 2X lysis buffer 1:1 with 18.2 MΩ water	Unadjusted stock solution at pH 8.5	1 year at room temp (RT)
2X lysis buffer (for liquid samples)	10% SDS, 100 mM TEAB pH 8.5	Unadjusted stock solution at pH 8.5	1 year at RT
Reductant	120 mM TCEP in water	Unadjusted	Prepare fresh
Alkylator	500 mM MMTS in isopropanol	Unadjusted	1 month at 4 °C
Acidifier, micros (27.5% phosphoric acid)	Phosphoric acid diluted to 27.5% with water. E.g. 324 µL of 85% phosphoric acid diluted into 676 µL of water.	Unadjusted	1 year at RT if kept sealed; can absorb CO ₂ from air over time
1 M TEAB stock*	1 M TEAB; pH adjusted with phosphoric acid	7.55	Freeze 1.5 mL aliquots in 15 mL falcon tubes
Binding/wash buffer*†	100 mM TEAB (final) in 90% methanol. Dilute the above TEAB stock with MeOH: e.g. to 1 mL 1 M TEAB, add MeOH until the final volume is 10 mL	7.55	1 month at RT; 1 year at 4 °C
Trypsin stock solution	Trypsin resuspended in 50 mM TEAB at a concentration of 1 mg/mL	Unadjusted pH at 8.5	Aliquot and freeze at -80 °C
Digestion buffer*†	50 mM TEAB containing sufficient trypsin stock solution yield 1 µg trypsin per 10 µg sample (weight:weight) in 20 µL 50 mM TEAB; no less than 1 µg trypsin per micro column	Unadjusted pH at 8.5	1 month at RT; 1 year at 4 °C
Elution buffer 1†	50 mM TEAB in water	Unadjusted pH at 8.5	1 month at RT; 1 year at 4 °C
Elution buffer 2	0.2% formic acid in water	Unadjusted	1 month at RT; 1 year at 4 °C
Elution buffer 3	50% acetonitrile in water	Unadjusted	1 month at RT; 1 year at 4 °C

* Tris can be used in place of TEAB at the same concentrations and pH values. Note that tris is not compatible with iTRAQ, TMT or other amine labeling chemistries.

† Ammonium bicarbonate cannot be used in the binding/wash buffer due to insufficient solubility in 90% MeOH. Ammonium bicarbonate can be used in place of TEAB in the digestion buffer and elution buffer 1.

S-Trap micro protocol



Solubilize in SDS, reduce, alkylate

1) Add 23 μL 1X lysis buffer to pelleted cells or 11.5 μL 2X lysis buffer to 11.5 μL liquid sample to yield a final volume of 23 μL containing from 1 – 100 μg of protein.

- For samples of different volumes, see Appendix A at end of protocol.
- SDS concentration must be $\geq 2\%$ for good recovery. Add SDS to any samples which do not contain it.
- High concentrations of SDS (up to 15%) do not affect S-Trap performance. If necessary, concentrate samples on a SpeedVac or lyophilizer.

2) Sonicate or vortex to lyse sample and dissolve proteins.

- For dilute protein samples which may have adsorbed onto tubes: **a)** add sufficient SDS so that the sample, once concentrated, will be 5% SDS final; **b)** sonicate the sample in a bath sonicator for 5 min to recover adsorbed protein; **c)** concentrate; and **d)** resuspend in 23 μL 50 mM TEAB pH 8.5. Larger volumes than 23 μL may be needed if the sample contains salts, urea or other involatile components.

3) If the sample is viscous, shear DNA thoroughly via sonication or treatment with Benzonase. This is essential as DNA can clog the S-Trap column.

Approach	Steps
PIXUL ultrasonication	Ultrasonicate for 5 min using 50 N pulses at 1 kHz with a 20 Hz burst rate.
Covaris ultrasonication	On a S220, ultrasonicate at 175 W using 200 cycles per burst and a 10% duty factor at 20 °C for 6 minutes.
Probe or horn sonicator	Settings depend on the power of the sonicator. Test power setting and length of sonication with a 1.5 mg/mL solution of salmon sperm DNA, sonicating until the viscosity is indistinguishable from an aqueous solution.
Benzonase	Add MgCl_2 to lysis buffer to yield a final concentration of 2 mM. Add 100 units Benzonase per 50 μg of protein and incubate at RT for 5 min.

4) As needed, clarify sample of debris by centrifugation (e.g. 8 min at 13,000 g and take the supernatant).

- **Note:** insoluble pellets may contain proteins of interest. If desired, insoluble sample can be processed as solids atop a separate S-Trap. Perform all following steps on the solid sample, and homogenize the insoluble sample as much as possible before sample processing.

5) Reduce: add 1 μL of of reductant (final concentration 5 mM TCEP) and incubate at 55 °C for 15 min.

6) Alkylate disulfides: add 1 μL of alkylator (final concentration 20 mM MMTS) and incubate at RT for 10 min.

7) Add 2.5 μL of acidifier, micros, to the 25 μL sample (final concentration $\sim 2.5\%$ phosphoric acid). Vortex.

- This step is essential to completely denature proteins and trap them efficiently.
- The pH will be ≤ 1 . If the sample pH is not ≤ 1 , add additional phosphoric acid to reach pH ≤ 1 .
- The final phosphoric acid concentration is different between S-Trap micros, and minis/midis.

Trap protein

8) Add 165 µL of binding/wash buffer to the sample and mix.

- Sufficient protein may make the colloidal protein particulate visible, giving a translucent appearance.
- All sample – including any resulting colloid – must be transferred to the S-Trap. Do not centrifuge at this point.

9) Place S-Trap micro in a 1.7 mL receiver tube for waste flow through. Apply sample to S-Trap column.

- Transfer all sample including anything insoluble into the S-Trap.
- No column preequilibration is necessary.
- Do not add more solution than will fit in the narrow “stem” of the spin column. For larger initial volumes, see Appendix A and load the column multiple times until the full volume has been bound.

10) Centrifuge the S-Trap column at 4,000 g for 30 sec to trap proteins.

- Visually confirm all sample has passed through the column; if not, centrifuge again until all solution has passed through.
- Negative pressure via a vacuum manifold or positive pressure from above can also be used.
- Clogged columns may be centrifuged as high as 15,000 g. See above notes on DNA sheering.

Clean protein

11) Add 150 µL binding/wash buffer; centrifuge at 4,000 g for 30 sec. Repeat 3 times and discard flow through as necessary.

- For best results, rotate the S-Trap micro units (like a screw or knob) 180 degrees between the centrifugations of binding and wash steps. This is especially important when using a fixed-angle rotor because the spin column does not experience homogenous flow. A mark on the outside edge during centrifugation makes it easy to track rotations.
- Additional wash(es) may be performed and should be performed if contamination is observed by mass spectrometry. You cannot over wash mammalian proteins with binding/wash buffer.
- If the sample has particularly hydrophobic contaminants, such as lipids in brain, bone marrow or adipose tissue, or wax in FFPE, bind protein, wash three times with 50% CHCl₃/50% MeOH, filling the entire column each time, then perform three washes with standard methanolic binding/wash buffer as normal.
- Vacuum or pressure may also be used at this step. S-Trap micros fit into a vacuum manifold with lure lock fittings. For large initial sample volumes, use of a vacuum manifold is highly recommended.
- After each centrifugation step, make sure that all added solution has gone through the S-Trap column. Centrifuge longer as needed. Unsheered DNA, highly viscous proteins (e.g. from mucosal tissues) or spin column overloading may necessitate significantly longer spin times.

12) Centrifuge S-Trap column at 4,000 g for 1 min to fully remove binding/wash buffer. Failure to do so may result in dripping.

13) Transfer S-Trap micro column to a clean 1.7 mL sample tube for the digestion.

Incubate and digest protein

14) Add 20 µL of digestion buffer containing protease, most commonly trypsin or a trypsin/lys-C mix, at a 1:10 weight to weight (wt:wt) ratio into the top of the S-Trap. By example, for 50 µg of protein to digest, apply 5 µg of trypsin dissolved in 20 µL.

- Do not apply less than 1 µg of trypsin for effective digestion.
- 50 mM TEAB, tris or ammonium bicarbonate are recommended as digestion buffers for trypsin or trypsin/lys-C mixes.
- Visually confirm no air bubbles are present atop the trap. Bubbles prevent the digestion buffer from entering the trap. If bubbles are present, flick the tube to remove them and/or centrifuge extremely briefly on a bench top centrifuge, returning any flow through to the top of the spin column.
- The S-Trap is hydrophilic and will absorb the digestion buffer. Careful observation shows that the applied digestion buffer “sinks” by a fraction of a mm when first applied. Do not damage the matrix with pipette tips.
- Other proteases than trypsin can be used, however they may require different digestion buffers, temperatures, pHes, metal cofactors, and/or times. S-Traps have been found to be compatible with Tryp-N, Lys-N, Lys-C, Arg-C, Glu-C,

chymotrypsin, elastase and pepsin. Trypsin/lys-C mixes generally perform better than trypsin alone. S-Traps can also be used for glycomics with PNGase F.

- Mass spec compatible detergents such as Rapigest™ are compatible with S-Trap sample digestion and for some sample types have been observed to aid in digestion. ProteaseMAX™ should not be used at elevated digestion temperatures due to accelerated autolysis. Trypsin/lys-C mixes generally perform better than trypsin alone.

15) Cap the S-Trap loosely to limit evaporative loss.

- Do not make an airtight seal with the cap. An airtight seal will force the solution out of the S-Trap during incubation.

16) Incubate for 1 – 2 hrs at 47 °C for trypsin or trypsin/lys-C, or overnight at 37 °C. Preferably use a water bath or stationary thermomixer.

- Ensure the entire column is exposed to heat. Do not shake.
- Preload dry incubators with beakers of water to saturate the air with water vapor and limit evaporation; let the beakers come to temperature before using the incubator. If using a thermomixer, fill unused wells with water and cap. If samples appear to be drying out despite a water-saturated atmosphere, add additional digestion buffer to compensate for evaporative loss.
- **As with all digestions, optimization of amount of protease, digestion time, buffer and temperature is sample- and enzyme-dependent.** Peptides all have their own unique digestion kinetics: optimize as necessary. See <https://pubs.acs.org/doi/abs/10.1021/pr100656u>.

Elute peptides for analysis

17) Add 40 µL of elution buffer 1 to the S-Trap then centrifuge (4,000 rcf, 1 min).

- Do not centrifuge the digestion through before applying elution buffer 1. Apply elution buffer 1 directly into the trap containing the digestion buffer that was incubated.

18) Add 40 µL of elution buffer 2 to the S-Trap then centrifuge (4,000 rcf, 1 min).

19) Add 40 µL of elution buffer 3 to the S-Trap then centrifuge (4,000 rcf, 1 min). This elution assists in recovery of hydrophobic peptides. Other organics may also be used as needed.

20) Pool eluted peptides, dry down and resuspend as desired (e.g. aqueous buffer A such as 5% acetonitrile, 0.1% formic acid for reverse phase chromatography or MALDI matrix).

Appendix A – Buffer additions based on initial sample volumes

- If the protein sample does not contain SDS, either add the same volume of 2X lysis buffer, add dry SDS or add SDS stock to a minimum of 2% SDS.
- The amount of reductant to add is always 1/23 of the initial sample volume. E.g. for 23 µL, add 1 µL.
- The amount of alkylator to add is always 1/23 of the initial sample volume. E.g. for 23 µL, add 1 µL.
- The amount of acidifier to add is always 1/10 of the initial sample volume, plus the added volumes of TCEP and MMTS. E.g. for 25 µL, add 2.5 µL.
- The amount of binding/wash buffer to add is 6X the total volume of sample (initial volume plus other reagents). E.g. for (25 µL + 2.5 µL) x 6 = 165 µL, add 165 µL.

Sample	1X lysis buffer (µL)	2X lysis buffer (µL)	Reductant (µL)	Alkylator (µL)	Acidifier (µL)	Binding/wash buffer (µL)
Pelleted cells	23	—	1	1	2.5	165
Protein pellet	23	—	1	1	2.5	165
11.5 µL liquid	—	11.5	1	1	2.5	165
15 µL liquid	—	15	1.3	1.3	3.3	215
30 µL liquid	—	30	2.6	2.6	6.5	430
100 µL liquid	—	100	8.7	8.7	21.7	1,435

Appendix B – Sample type recommendations

Note that protein concentration should be assayed by BCA for all samples except IPs.

Sample type	Recommendation
Adherent cells on plate	Wash cell plate 3x with cold PBS in rapid succession ensuring complete PBS rinse removal each time. Do not let cells dry out. For best results, tip the plate containing the last PBS wash and use a cell scraper to bring all cells to the bottom corner of the plate. Remove cells, pellet and lyse with 1X lysis buffer using sonication.
Bacteria and archaea	Disrupt the cell walls using beads beating, cryopulverization and/or ultrasonication in 1X lysis buffer. Degree of homogenization needed is strongly a function of the species of microorganism. Monitor by microscopy. See https://doi.org/10.3390/microorganisms8030413 .
Bile	Mix bile 1:1 with 2X lysis buffer and proceed with S-Trap digestion. See https://doi.org/10.1016/j.jprot.2020.103984 .
Bioreactor supernatant	Mix bioreactor supernatant 1:1 with 2X lysis buffer and proceed with the S-Trap digestion. Polymeric surfactants such as Pluronic F68 will be fully removed. See https://doi.org/10.1021/acs.jproteome.0c00106 .
Exosomes	Purify exosomes by differential centrifugation. Dissolve exosomes in 1X lysis buffer preferably with ultrasonication. Proceed with S-Trap digestion. See https://doi.org/10.1038/s41598-018-37002-x .
Formalin-fixed paraffin-embedded (FFPE) samples, scrolls or cores	Follow the HYPERsol protocol (https://doi.org/10.1021/acs.jproteome.9b00686) with the following alteration: allow cores or scrolls to hydrate in 1X lysis buffer overnight. The next day, ultrasonicate for 5 min or until pieces have been dissolved. For PIXUL ultrasonication, use the following settings: 50 N pulse at 1 kHz with a 20 Hz burst rate. Perform additional washes with 50% CHCl ₃ /50% MeOH as described above in “Clean protein.”
IP	Elute directly with 1X SDS lysis buffer (5% SDS) and use the high recovery protocol.
Membrane fractions/proteins	Isolate membrane fraction; ultracentrifugation or Triton X-114 phase separation can be used. Dissolve membrane fraction in 1X lysis buffer with ultrasonication and proceed with S-Trap digestion. See https://doi.org/10.1038/s41594-020-0425-5 .
Laser capture microdissection (LCM) pieces	Dissolve LCM pieces in 1X lysis buffer with heating and use the high recovery S-Trap protocol. See https://doi.org/10.1186/s12014-020-09287-6 . Note this reference analyzed haematoxylin and eosin (H&E) stained, formalin-fixed paraffin-embedded (FFPE) tissues.
Nonadherent cells	Gently pellet cells and resuspend in the maximal practical volume of cold PBS (e.g. ~10 mL in a 15 mL Falcon tube). Repeat rinse 3x, then pellet and lyse with 1X lysis buffer using ultrasonication.
PBMC fractions, COVID samples	Lyse cells or sample in 1X lysis buffer and proceed with the S-Trap digestion. See https://doi.org/10.1021/acs.jproteome.0c00365 .
Saliva or sputum	Measure protein concentration and aliquot saliva sufficient for 100 µg. Add a 20% stock solution of SDS to 2% SDS. Reduce and alkylate to liquefy, then follow the standard S-Trap protocol. See https://doi.org/10.1021/acs.jproteome.8b00505 .
Serum/plasma	Dilute serum/plasma 1:20 with PBS and mix 12.5 µL diluted serum with 12.5 µL 2X lysis buffer. See https://doi.org/10.1016/j.jprot.2020.103645 .
Stool	Dissolve stool with 1X lysis buffer using agitation (bead beating or ultrasonication) and proceed with S-Trap digestion. See https://doi.org/10.1128/mSystems.00200-20 .
Tissues	Do not allow tissue to defrost until immediately before lysis. As needed and depending on size, crush frozen tissue with a precooled pellet pestle, mortar and pestle, hammer, Covaris CryoPrep, etc. In all cases be extremely cautious to not lose pieces of tissue as THEY WILL FLY OUT . Ideally, cryopulverize and/or ultrasonication in 1X lysis buffer. Add frozen 1X lysis buffer at 10X – 50X wt:wt with tissue. The degree of homogenization needed is strongly a function of the kind of tissue. Bone will require demineralization with EDTA; larger pieces may require days. Length and power of ultrasonication must be optimized for each tissue.
Urine	Aliquot 0.5 mL urine containing at least 100 µg of protein and add 2 mL of cold acetone. Precipitate the proteins overnight at –20 °C. Centrifuged samples at 20,000 g for 10 min and dissolve the pellet in lysis buffer. See https://doi.org/10.1021/acs.jproteome.9b00772 .
Yeast	Disrupt the cell walls using beads beating, cryopulverization, heat and/or ultrasonication in 1X lysis buffer. Monitor cell wall disruption by microscopy.

Appendix C – Troubleshooting

In general, troubleshoot protein capture and digestion by gel analysis of the flow through, washes and elutions. An additional SDS strip (for gel analysis) may be used to determine absolute recovery.

Problem	Possible causes and solutions
Protein not captured or is “missing”	<ol style="list-style-type: none"> 1. Applied protein sample did not contain sufficient SDS. SDS is necessary for efficient capture in the protein-trapping matrix. Make sure applied sample contains 5% SDS. 2. Applied protein sample was not acidified properly with phosphoric acid. This step is also necessary for efficient protein denaturation and capture. Make sure the SDS solubilized protein sample is acidified to a final concentration of 2.5% phosphoric acid and is highly acidic. Check pH is ≤ 1. 3. All protein, including any particulate or colloid, must be transferred into the S-Trap unit. If the acidified SDS lysate/MeOH S-Trap buffer solution was centrifuged before addition into the unit, it is possible the protein was pelleted out and did not enter the unit. Ensure complete transfer of acidified lysate/S-Trap buffer.
Incomplete protein digestion	<ol style="list-style-type: none"> 1. Protease concentration may need to be optimized depending on sample and protease. Try different weight:weight concentrations of protease; try different digestion temperatures; include a mass spec compatible detergent like Rapigest™ or ProteaseMAX™; and/or increase digestion time. Due to evaporation, digestion volume will likely need to be increased with digestion times longer than 1 hr. Do not apply less than 1 μg of trypsin per trap for efficient digestion; other proteases may require different minimum amounts. Trypin/Lys-C mixes work better than trypsin alone. 2. Heating of the S-Trap may be insufficient. Ensure entire S-Trap column is exposed to heat and <u>do not</u> have the S-Trap column e.g. sticking out of the top of a heat block where it will receive insufficient heating. 3. Bubbles atop the trap may prevent the sample from being exposed to the digestion protease; ensure no bubbles are present by visual inspection. If any bubbles are present, flick until they float to the top. Protease may be applied directly above the trap with a thin pipette tip. Note that the protein trap is highly hydrophilic; the protease solution will be quickly “sucked” into the trap in the absence of a bubble. With careful observation, this behavior can be seen – it lasts < 1 sec – and is verification that the protease has been absorbed into the trap. 4. See 3 in “protein not captured or is ‘missing’”: especially if using a fixed-angle centrifuge, protein may have been pelleted outside the protein-trapping matrix, for example on the side of the S-Trap column, and not exposed to protease. Rotate the S-Trap column 180 degrees as specified in step 11; alternatively use a swing-bucket rotor or a vacuum manifold.
Poor peptide recovery	<ol style="list-style-type: none"> 1. The size of S-Trap unit should be matched to the amount of protein to digest. S-Trap micros are not recommended for > 100 μg; use S-Trap minis. MS-compatible detergents have been found to aid in recovery of low protein levels. Additional organic elutions may also aid in recovery. 2. If the digest has dried on the column (for example, if the cap was not applied during incubation, or the column was forgotten), the protein-trapping matrix will need to be rehydrated to solubilize the peptides. Add digestion buffer without enzyme and let sit for 30 min, then centrifuge out. Repeat the wash and elution of steps. Additional elutions may assist in peptide recovery. Concentrate by lyophilization. 3. The protein-trapping matrix retains proteins but not peptides. If digestion is incomplete (see above), poor peptide recovery will result. Digestion can be monitored by gel as well as mass spec analyses.